



# Same-Day Identification and Antimicrobial Susceptibility Testing of Bacteria in Positive Blood Culture Broths Using Short-Term Incubation on Solid Medium with the MicroFlex LT, Vitek-MS, and Vitek2 Systems

Jihye Ha, M.D.<sup>1</sup>, Sung Kuk Hong, M.D.<sup>2</sup>, Geum Hee Han, M.T.<sup>1</sup>, Myungsook Kim, M.T.<sup>1</sup>, Dongeun Yong, M.D.<sup>1</sup>, and Kyungwon Lee, M.D.<sup>1</sup>

Department of Laboratory Medicine<sup>1</sup> and Research Institute of Bacterial Resistance, Yonsei University College of Medicine, Seoul; Department of Laboratory Medicine<sup>2</sup>, National Cancer Center, Goyang, Korea

**Background:** Early and appropriate antibiotic treatment improves the clinical outcome of patients with septicemia; therefore, reducing the turn-around time for identification (ID) and antimicrobial susceptibility test (AST) results is essential. We established a method for rapid ID and AST using short-term incubation of positive blood culture broth samples on solid media, and evaluated its performance relative to that of the conventional method using two rapid ID systems and a rapid AST method.

**Methods:** A total of 254 mono-microbial samples were included. Positive blood culture samples were incubated on blood agar plates for six hours and identified by the MicroFlex LT (Bruker Daltonics) and Vitek-MS (bioMérieux) systems, followed by AST using the Vitek2 System (bioMérieux).

**Results:** The correct species-level ID rates were 82.3% (209/254) and 78.3% (199/254) for the MicroFlex LT and Vitek-MS platforms, respectively. For the 1,174 microorganism/antimicrobial agent combinations tested, the rapid AST method showed total concordance of 97.8% (1,148/1,174) with the conventional method, with a very major error rate of 0.5%, major error rate of 0.7%, and minor error rate of 1.0%.

**Conclusions:** Routine implementation of this short-term incubation method could provide ID results on the day of blood culture-positivity detection and one day earlier than the conventional AST method. This simple method will be very useful for rapid ID and AST of bacteria from positive blood culture bottles in routine clinical practice.

**Key Words:** Antimicrobial susceptibility testing, Rapid identification, Septicemia, MALDI-TOF-MS, Blood culture

**Received:** March 24, 2017

**Revision received:** August 16, 2017

**Accepted:** December 14, 2017

**Corresponding author:** Dongeun Yong  
Department of Laboratory Medicine, Yonsei University College of Medicine, 50 Yonsei-ro, Seodaemun-gu, Seoul 03722, Korea  
Tel: +82-2-2228-2442  
Fax: +82-2-364-1583  
E-mail: deyong@yuhs.ac

**Co-corresponding author:** Sung Kuk Hong  
Department of Laboratory Medicine, National Cancer Center, Goyang, 323 Ilsan-ro, Ilsandong-gu, Goyang 10408, Korea  
Tel: +82-31-920-1324  
Fax: +82-2-364-1583  
E-mail: skhong@ncc.re.kr

## © Korean Society for Laboratory Medicine

This is an Open Access article distributed under the terms of the Creative Commons Attribution Non-Commercial License (<http://creativecommons.org/licenses/by-nc/4.0>) which permits unrestricted non-commercial use, distribution, and reproduction in any medium, provided the original work is properly cited.

## INTRODUCTION

Appropriate and timely antimicrobial treatment significantly improves clinical outcomes for patients with sepsis [1]. Therefore,

reducing the turn-around time of species identification (ID) and antimicrobial susceptibility test (AST) results is a high-priority topic for clinical microbiologists. The introduction of matrix-assisted laser desorption ionization-time of flight mass spectrometry

try (MALDI-TOF MS) has allowed pathogen species ID within a few minutes, as opposed to the several days required by conventional biochemical methods [2]. Unlike other clinical samples, positive blood cultures need at least one additional day to yield pure bacterial colonies for subsequent tests and procedures. Therefore, various blood sample preparation protocols for MALDI-TOF MS have been developed to remove human cells and proteins from the blood culture broth [3-7]. However, these sample preparation methods negate the main advantages of this technique—convenience and speed—by introducing additional labor-intensive and expensive steps, and this drawback has limited their application in routine clinical microbiology practice.

To overcome the drawbacks of the current sample preparation methods, we developed an easier and cheaper method, designated the “short-term incubation method on a solid medium.” This method could eliminate the need for centrifugation and additional chemicals to extract bacterial proteins. This protocol involves a 6-hour incubation of a positive blood culture broth with subsequent ID and AST using MALDI-TOF MS and automated AST devices. This is the first prospective study to evaluate the performance of this short-term incubation method for both species ID and AST. This study also compares the performance of two rapid ID platforms using this approach: MicroFlex LT (Bruker Daltonics, Bremen, Germany) and Vitek-MS (bioMérieux, Marcy l’Etoile, France).

## METHODS

### 1. Blood culture samples

We used blood culture bottles inoculated with blood drawn from a peripheral vein that were referred to the laboratory in Sever-

ance hospital, Seoul, Korea, during the period from April to August 2014. Any sample with growth of more than one bacterial species or yeast was excluded from the study. Among the 334 blood cultures referred, 254 mono-microbial aerobic samples were ultimately identified and included in the study. The BacT/ALERT® 3D system (bioMérieux, Marcy l’Etoile, France) was used for initial testing of the blood cultures. When a blood culture bottle showed a positive signal, Gram staining was performed.

Parallel to the conventional workflow shown in Fig. 1, broth was collected from the positive blood culture bottles and inoculated on blood agar plates (BAP; Asan Pharmaceutical Co., Ltd., Seoul, Korea) and MacConkey agar plates (Becton Dickinson, Sparks, MD, USA), followed by incubation at 35°C under 5% CO<sub>2</sub> atmosphere. One BAP was used for the rapid ID and AST, and the others were used for conventional biochemical ID and AST. This study was approved by the Institutional Review Board of Severance Hospital, Yonsei University Health System (2017-2648-001).

### 2. Conventional workflow of positive blood cultures

When the BacT/ALERT® 3D System showed a positive signal, Gram staining was performed, followed by subculture on an appropriate solid agar medium. Following overnight incubation, the colonies grown on the agar plates were used for ID and AST using the commercial automated Vitek2 system (bioMérieux). The ID and AST results obtained using this conventional workflow were used as the standard for comparison.

### 3. Rapid ID and AST using the short-term incubation method

Rapid ID was performed using two MALDI-TOF MS systems: MicroFlex LT and Vitek-MS. The Vitek2 system was used for the




Fig. 1. Workflow of rapid and conventional species identification (ID) and antimicrobial susceptibility test (AST).

rapid AST.

Bacterial colonies that grew on the BAP after the 6-hour incubation period were transferred to the MicroFlex LT and Vitek-MS target plates using a 1- $\mu$ L sterile plastic loop. The target plates were overlaid with 1  $\mu$ L of a matrix solution composed of saturated  $\alpha$ -cyano-4-hydroxycinnamic acid in 50% acetonitrile/2.5% trifluoroacetic acid or with the VitekMS-CHCA matrix solution ( $\alpha$ -cyano-4-hydroxycinnamic acid solution, ethanol, acetonitrile, solvent) for MicroFlex LT or Vitek-MS, respectively. The plates were dried in room air and subsequently subjected to MALDI-TOF MS. The ID process was performed only once for each sample. Criteria for reliable ID included confidence scores of  $\geq 1.7$  for MicroFlex LT and  $\geq 90\%$  for Vitek-MS [8, 9] according to the manufacturers' guidelines. When Vitek-MS provided more than one result with the same genus, we assumed that the result represented a reliable ID at the genus level.

To compare the AST results obtained using the rapid AST and standard methods, the minimum inhibitory concentrations obtained by both methods were categorized as susceptible, intermediate, or resistant, according to the interpretive criteria of the Vitek2 system following the CLSI recommendation [10]. The comparison between the rapid and standard methods was classified as: agreement, very major error (false susceptibility), major error (false resistance), or minor error (susceptible/resistant vs intermediate susceptibility). AST was performed for 232 isolates, which could be possible pathogens considering the patients' clinical characteristics.

#### 4. Time to ID

We recorded the turn-around time for both the conventional and rapid ID methods from the time of sample registration to result reporting through the five-month study duration from April to

**Table 1.** Results obtained by the two MALDI-TOF MS systems using the short-term bacterial incubation method compared with those of the conventional method

Microorganism	N	N (%) identified in MicroFlex			N (%) identified in Vitek-MS		
		Concordance level		Non-reliable ID	Concordance level		Non-reliable ID
		Species	Genus		Species	Genus	
Gram-positive bacteria	143	110 (76.9)	116 (81.1)	28 (19.6)	105 (73.4)	110 (76.9)	33 (23.1)
Gram-positive cocci	120	100 (83.3)	106 (88.3)	15 (12.5)	96 (80.0)	101 (84.2)	19 (15.8)
<i>Staphylococcus aureus</i>	14	14 (100)	14 (100)	0 (0.0)	14 (100)	14 (100)	0 (0.0)
Coagulase-negative staphylococci*	60	52 (86.7)	53 (88.3)	7 (11.7)	47 (78.3)	48 (80.0)	11 (18.3)
Enterococci	25	24 (96.0)	24 (96.0)	1 (4.0)	24 (96.0)	24 (96.0)	1 (4.0)
<i>Enterococcus faecalis</i>	12	12 (100)	12 (100)	0 (0.0)	12 (100)	12 (100)	0 (0.0)
<i>Enterococcus faecium</i>	13	12 (92.3)	12 (92.3)	1 (7.6)	12 (92.3)	12 (92.3)	1 (7.6)
Streptococci <sup>†</sup>	19	10 (52.6)	14 (73.7)	7 (36.8)	11 (57.9)	14 (73.7)	6 (31.6)
Other <sup>‡</sup>	2	1 (50.0)	1 (50.0)	1 (50.0)	1 (50.0)	1 (50.0)	1 (50.0)
Gram-positive rods <sup>§</sup>	23	10 (43.5)	10 (43.5)	13 (56.5)	9 (39.1)	9 (39.1)	14 (60.9)
Gram-negative bacteria	111	99 (89.2)	106 (95.5)	5 (4.5)	94 (84.7)	100 (90.9)	11 (9.9)
Enterobacteriaceae	94	85 (90.4)	90 (95.7)	4 (4.3)	79 (84.0)	85 (90.4)	9 (9.6)
<i>Escherichia coli</i>	53	52 (98.1)	52 (98.1)	1 (1.9)	51 (96.2)	51 (96.2)	2 (3.8)
<i>Klebsiella pneumoniae</i>	22	20 (90.9)	20 (90.9)	2 (9.1)	17 (77.3)	17 (77.3)	5 (22.7)
Other Enterobacteriaceae <sup>  </sup>	19	13 (68.4)	18 (94.7)	1 (5.3)	11 (57.8)	17 (89.4)	2 (10.5)
Glucose non-fermenters <sup>¶</sup>	11	9 (81.8)	10 (90.9)	1 (9.1)	10 (90.9)	10 (90.9)	1 (9.1)
Other Gram-negative bacteria**	6	5 (83.3)	6 (100)	0 (0.0)	5 (83.3)	5 (83.3)	1 (16.7)
Total	254	209 (82.3)	222 (87.4)	33 (13.0)	199 (78.3)	210 (86.6)	44 (17.3)

\*Coagulase-negative staphylococci: *S. epidermidis*, *S. hominis*, *S. haemolyticus*, *S. capitis*, *S. auricularis*, *S. saprophyticus*, *S. warneri*; <sup>†</sup>Streptococci: *S. pyogenes*, *S. agalactiae*, *S. anginosus*, *S. constellatus*, *S. intermedius*, *S. salivarius*, *S. pluranimalium*, *S. equisimilis*, *S. pneumoniae*, *S. mitis*, *S. parasanguinis*, *S. sanguis*; <sup>‡</sup>Other: *Granulicatella adiacens*; <sup>§</sup>Gram-positive rods: *Bacillus* spp., diphtheroids, *Clostridium septicum*, *Arthrobacter* spp., *Listeria monocytogenes*; <sup>||</sup>Other Enterobacteriaceae: *Enterobacter cloacae*, *Enterobacter asburiae*, *Enterobacter aerogenes*, *Citrobacter freundii*, *Citrobacter braakii*, *Morganella morganii*, *Salmonella* spp., *Proteus mirabilis*; <sup>¶</sup>Glucose non-fermenters: *Acinetobacter baumannii*, *Acinetobacter lwoffii*, *Acinetobacter radioresistens*, *Pseudomonas aeruginosa*, *Stenotrophomonas maltophilia*; \*\*Other Gram-negative bacteria: *Burkholderia cepacia*, *Aeromonas caviae*, *Bacteroides caccae*, *Mycobacterium abscessus*, *Ochrobactrum anthropicum*.

August 2014. The mean difference was calculated based on records from the laboratory information system.

## 5. Statistical analysis

Chi-square test was used to compare proportions. Statistical analyses were performed using a software package (Analyse-it, version 3.90.7). A *P* value less than 0.05 was considered statistically significant.

## RESULTS

### 1. Comparison of rapid and conventional ID results

The concordance of the species-level ID results obtained using MicroFlex LT and Vitek-MS with those obtained using the con-

ventional biochemical ID method (Vitek2 system) was 82.3% (209/254) and 78.3% (199/254), respectively (Table 1, see Supplemental Data Table S1). The concordance rates for Gram positive bacteria were 76.9% (110/143) and 73.4% (105/143) and the rates for Gram-negative bacteria were 89.2% (99/111) and 84.7% (94/111) for the MicroFlex LT and Vitek-MS systems, respectively.

Table 1 also shows the percentages of non-reliable results, which were slightly higher for the Vitek-MS system than for the MicroFlex system, but the difference was not statistically significant ( $P=0.174$ ). In addition, the non-reliable ID percentage was greater for Gram-positive isolates than for Gram-negative bacteria with both systems. Considering only the reliable ID results, the species-level concordance rates were more than 94% and

**Table 2.** Antimicrobial susceptibility test results for the short-term incubation method compared with those of the conventional method

Microorganism (N of ASTs conducted)	N (%) of			
	Agreement	Minor error	Major error	Very major error
Gram-positive bacteria (517)	509 (98.5)	3 (0.6)	3 (0.6)	2 (0.4)
Staphylococci (368)	364 (98.9)			
<i>S. aureus</i> (80)	80 (100)			
<i>S. epidermidis</i> (128)	126 (98.4)	One with teicoplanin, one with ciprofloxacin		
<i>S. hominis</i> (96)	94 (97.9)		One with cotrimoxazole	One with cotrimoxazole
<i>S. haemolyticus</i> (16)	16 (100)			
<i>S. capitis</i> (32)	32 (100)			
<i>S. auricularis</i> (16)	16 (100)			
Enterococci (806)	785 (97.4)			
<i>E. faecium</i> (65)	64 (98.5)	One with nitrofurantoin		
<i>E. faecalis</i> (84)	81 (96.4)		One with penicillin G, one with gentamicin	One with vancomycin
Gram-negative bacteria (657)	639 (97.3)	9 (1.4)	5 (0.8)	4 (0.6)
Enterobacteriaceae (608)	590 (97.0)			
<i>Escherichia coli</i> (288)	275 (95.5)	Five with cefepime, one with ampicillin-sulbactam, one with meropenem	One with ampicillin-sulbactam, one with gentamicin, one with cotrimoxazole, one with aztreonam	One with ceftazidime, one with ertapenem
<i>Klebsiella pneumoniae</i> (160)	159 (99.4)			One with aztreonam
<i>Enterobacter cloacae</i> (112)	109 (100)	Two with ertapenem		One with meropenem
<i>Enterobacter asburiae</i> (16)	16 (100)			
<i>Citrobacter freundii</i> (16)	16 (100)			
<i>Citrobacter braakii</i> (16)	15 (93.8)		One with amikacin	
Glucose non-fermenters (49)	49 (100)			
<i>Acinetobacter baumannii</i> (16)	16 (100)			
<i>Acinetobacter radioresistens</i> (17)	17 (100)			
<i>Pseudomonas aeruginosa</i> (16)	16 (100)			
Total (1,174)	1,148 (97.8)	12 (1.0)	8 (0.7)	6 (0.5)

Abbreviation: AST, antimicrobial susceptibility test.

were similar for both systems. However, the concordance rates for streptococci (n = 19) were considerably low for both systems.

## 2. Comparison of rapid and conventional AST results

Among the 254 isolates from the positive blood culture bottles, the AST results of 232 isolates showed clinical significance and were included in the analysis. As shown in Table 2, the rapid AST results exhibited very high concordance (>97%) with those of the conventional method, with a low error rate (<1.0%). The concordance rate was slightly higher for Gram-positive isolates (n=517) than for Gram-negative isolates (n=657), and the error rates were slightly lower in the former group. Interestingly, errors in *Escherichia coli* AST accounted for 72.2% (13/18) of the total error observed for Gram-negative isolates (Table 2).

## 3. Time to identification

The mean turn-around time for the conventional method over the five-month period was 5,691 minutes (3.95 days), whereas the rapid ID approach using the new short-term incubation method produced results in 3,722 minutes (2.58 days), representing a mean difference of 1,366 minutes (1.37 days).

## DISCUSSION

The aim of the present study was to evaluate the performance of a short-term incubation method for identifying bacterial pathogens from blood culture for both ID and AST, and to compare its results with those by the MicroFlex LT and Vitek-MS systems.

Bacteremia results in a high crude (24%) and attributable (17%) mortality rate [11]; therefore, appropriate and timely antimicrobial treatment is crucial for improving the clinical outcomes of the patients with bacteremia [1]. To eliminate the time-consuming step of conventional blood culture processing, several approaches, including MALDI-TOF MS, have been evaluated for achieving rapid pathogen ID from a positive blood culture broth [12, 13]. However, the complicated sample handling steps have hampered the widespread usage of MALDI-TOF MS in routine blood culture practice. To overcome this limitation, several methods for preparation of blood culture broth have been introduced to enable the direct identification of bacteria in positive blood culture bottles, such as lysis-filtration [3], lysis extraction [4], protein extraction [5], red blood cell lysis [6], and modified lysis with nylon mesh [7]. These methods could provide a positive ID for 91.4–99% of Gram-negative isolates and for 67.7–97% of Gram-positive isolates in about 15–45 minutes. However, all these methods require special reagents and a centrifugation

process. In the short-term incubation method developed in this study, the additional preparatory steps are minimized to the preparation of one BAP for one drop of a sample.

Although this proposed rapid ID and AST method has proven to be a very powerful and simple tool, there are some drawbacks that need to be overcome [14, 15]. According to Kohlmann *et al* [14], rapid ID results obtained using this method allowed for optimized treatment to be recommended in 51.1% of cases, whereas only 26.4% of the Gram stain results allowed for proper treatment. Zebbe *et al* [15] reported that 77% of bacteria were correctly identified with only a three-hour incubation of one drop from a positive blood culture bottle on a chocolate agar plate. Similar results were obtained in our study, in that the concordance rate was higher in Gram-negative isolates than in Gram-positive isolates. The higher percentage of non-reliable IDs in Gram-positive isolates could have resulted from the relatively less growth of these organisms after the six-hour incubation.

We incubated the subculture plates for six hours because this was previously shown to substantially increase the species-level ID concordance rate of the short-term incubation method to 90.9% (159/175) from 80.6% (141/175) obtained using a four-hour incubation period [16]. One study showed that the incubation time could be shortened by 3–6 hours using pre-warmed chocolate agar to obtain a similar correct identification rate [17]. However, that study used a pellet streak, which requires an additional centrifugation step to obtain the pellet from the blood culture broths. We did not adopt this process because this could represent a hurdle for routine practice.

Interestingly, most of the results that were discordant between MALDI-TOF MS and the conventional ID method were due to non-reliable results rather than misidentifications. Therefore, when samples with non-reliable results were excluded, very high concordance rates for species-level ID were obtained: 94.6% (209/221) for MicroFlex LT and 94.8% (199/210) for Vitek-MS. Surprisingly, low concordance rates of 52.6% (10/19) and 57.9% (11/19) were obtained for *Streptococcus* spp. with the MicroFlex LT and Vitek-MS platforms, respectively. This finding is in contrast with previous data, as summarized in a review article [18]. The cause of this discrepancy might be the difference in incubation times or, possibly, the lack of optimal conditions for the growth of streptococci on BAPs, which could result in a failure to produce sufficient proteins for accurate MALDI-TOF MS analysis. However, given the small number of streptococci isolates in this study, the results for streptococcal species identification might be considered preliminary when specifically using the short-term incubation method on a solid medium presented



herein.

Regarding the applications of this rapid procedure for AST, the error rate was not substantial, with 1.0% minor error, 0.7% major error, and 0.5% very major error; the results were within the <3% and <1.5% US Food and Drug Administration (FDA) limits for major and very major errors, respectively [19]. The AST results for *E. coli* showed the largest number of discordances, which is likely due to the fact that the number of antimicrobials tested in *E. coli* (average, 275) was much higher than that tested for the other species (average, 68). Despite this higher rate of discordance, the major and very major error rates (1.4% [4/288] and 0.7% [2/288], respectively) for *E. coli* were still within the FDA limits. According to the comparison of all AST results, the concordance rates between the conventional AST methods and the short-term incubation method presented herein were very high.

One possible limitation of this study is the restricted spectrum of bacterial species analyzed, which was due to the limited sample size obtained from a clinical microbiology laboratory. Thus, these results might not apply to other laboratories. Further study using a larger number of more diverse samples over a longer period is therefore warranted to validate these results.

In conclusion, the short-term incubation on solid medium method for positive blood cultures shows acceptable performance for routine ID and AST, especially in clinically relevant bacteria. This simple and rapid method could facilitate the implementation of rapid ID and AST into the routine workflow of clinical microbiology laboratories, which could improve patient outcomes.

## Authors' Disclosures of Potential Conflicts of Interest

No potential conflicts of interest relevant to this article were reported.

## Acknowledgments

This work was supported by Nano Material Technology Development Program through the National Research Foundation of Korea (NRF) funded by the Ministry of Science and ICT (No.2017 M3A7B4039936), by the Korea Health Technology R&D Project through the Korea Health Industry Development Institute (KH-IDI), funded by the Ministry of Health & Welfare, Republic of Korea (grant number HI17C1807), and by the BioNano Health-Guard Research Center funded by the Ministry of Science, ICT

& Future Planning (MSIP) of Korea as a Global Frontier Project (H-GUARD\_2014M3A6B2060509).

## REFERENCES

1. Paul M, Shani V, Muchtar E, Kariv G, Robenshtok E, Leibovici L. Systematic review and meta-analysis of the efficacy of appropriate empiric antibiotic therapy for sepsis. *Antimicrob Agents Chemother* 2010;54:4851-63.
2. Seng P, Rolain JM, Fournier PE, La Scola B, Drancourt M, Raoult D. MALDI-TOF-mass spectrometry applications in clinical microbiology. *Future Microbiol* 2010;5:1733-54.
3. Machen A, Drake T, Wang YF. Same day identification and full panel antimicrobial susceptibility testing of bacteria from positive blood culture bottles made possible by a combined lysis-filtration method with MALDI-TOF VITEK mass spectrometry and the VITEK2 system. *PLoS One* 2014;9:e87870.
4. Prod'homme G, Bizzini A, Durussel C, Bille J, Greub G. Matrix-assisted laser desorption ionization-time of flight mass spectrometry for direct bacterial identification from positive blood culture pellets. *J Clin Microbiol* 2010;48:1481-3.
5. Monteiro J, Inoue FM, Lobo AP, Sugawara EK, Boaretti FM, Tufik S. Fast and reliable bacterial identification direct from positive blood culture using a new TFA sample preparation protocol and the Vitek<sup>®</sup> MS system. *J Microbiol Methods* 2015;109:157-9.
6. Jakovljevic A and Bergh K. Development of a rapid and simplified protocol for direct bacterial identification from positive blood cultures by using matrix-assisted laser desorption ionization time-of-flight mass spectrometry. *BMC Microbiol* 2015;15:258.
7. Jo SJ, Park KG, Han K, Park DJ, Park YJ. Direct identification and antimicrobial susceptibility testing of bacteria from positive blood culture bottles by matrix-assisted laser desorption/ionization time-of-flight mass spectrometry and the Vitek2 system. *Ann Lab Med* 2016;36:117-23.
8. Chen JH, Ho PL, Kwan GS, She KK, Siu GK, Cheng VC, et al. Direct bacterial identification in positive blood cultures by use of two commercial matrix-assisted laser desorption ionization-time of flight mass spectrometry systems. *J Clin Microbiol* 2013;51:1733-9.
9. Buchan BW, Riebe KM, Ledebor NA. Comparison of the MALDI Biotyper system using Sepsityper sample processing to routine microbiological methods for identification of bacteria from positive blood culture bottles. *J Clin Microbiol* 2012;50:346-52.
10. CLSI. Performance standards for antimicrobial susceptibility testing. 24th ed. CLSI supplement M100-S24. Wayne, PA: Clinical and Laboratory Standards Institute; 2014.
11. Diekema DJ, Beekmann SE, Chapin KC, Morel KA, Munson E, Doern GV. Epidemiology and outcome of nosocomial and community-onset bloodstream infection. *J Clin Microbiol* 2003;41:3655-60.
12. Huang AM, Newton D, Kunapuli A, Gandhi TN, Washer LL, Isip J, et al. Impact of rapid organism identification via matrix-assisted laser desorption/ionization time-of-flight combined with antimicrobial stewardship team intervention in adult patients with bacteremia and candidemia. *Clin Infect Dis* 2013;57:1237-45.
13. Vlek AL, Bonten MJ, Boel CH. Direct matrix-assisted laser desorption ionization time-of-flight mass spectrometry improves appropriateness of antibiotic treatment of bacteremia. *PLoS One* 2012;7:e32589.
14. Kohlmann R, Hoffmann A, Geis G, Gatermann S. MALDI-TOF mass spectrometry following short incubation on a solid medium is a valuable tool for rapid pathogen identification from positive blood cultures. *Int J*

- Med Microbiol 2015;305:469-79.
15. Zabbe JB, Zanardo L, Mégraud F, Bessède E. MALDI-TOF mass spectrometry for early identification of bacteria grown in blood culture bottles. *J Microbiol Methods* 2015;115:45-6.
  16. Hong SK, Chang BK, Song SH, Kim EC. Use of MALDI-TOF MS technique for rapid identification of bacteria from positive blood cultures. *Indian J Med Microbiol* 2014;32:419-22.
  17. Bhatti MM, Boonlayangoor S, Beavis KG, Tesic V. Rapid identification of positive blood cultures by matrix-assisted laser desorption ionization-time of flight mass spectrometry using prewarmed agar plates. *J Clin Microbiol* 2014;52:4334-8.
  18. Fan WT, Qin TT, Bi RR, Kang HQ, Ma P, Gu B. Performance of the matrix-assisted laser desorption ionization time-of-flight mass spectrometry system for rapid identification of streptococci: a review. *Eur J Clin Microbiol Infect Dis* 2017;36:1005-12.
  19. Center for Devices and Radiological Health, Food and Drug Administration. Class II Special Controls Guidance Document: Antimicrobial Susceptibility Test (AST) Systems, <https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm388961.pdf> (Updated on Aug 2009).