

Erratum: Measurement and Simulation of Stilbene Scintillator Response for the KSTAR Neutron Diagnostic System

[J. Korean Phys. Soc. **64**, 645 (2014)]

Seung Kyu LEE

*Korea Atomic Energy Research Institute, 989-111 Daedeok-daero, Yuseong, Daejeon 305-353, Korea,
and Department of Nuclear Engineering, Hanyang University, Seoul 133-791, Korea*

Jae Bum SON and Jeong Soo KANG

Department of Nuclear Engineering, Hanyang University, Seoul 133-791, Korea

Hee SEO, Byung-Hee WON, Se-Hwan PARK and Ho-Dong KIM

Korea Atomic Energy Research Institute, Daejeon 305-353, Korea

Byoung Hwi KANG and Gi Dong KIM

Institute for Basic Science, Daejeon 305-811, Korea

Yong Kyun KIM*

*Department of Nuclear Engineering, Hanyang University, Seoul 133-791, Korea,
and Institute for Basic Science, Daejeon 305-811, Korea*

DOI: 10.3938/jkps.64.1234

Table 1 should be included in this paper by the following table.

Table 1. General properties of the stilbene scintillator.

Property	Value
Light output (% anthracene)	50
Density (g/cm ³)	1.22
H/C ratio	0.857
Decay time (ns)	3.5
Luminescence max (nm)	390
Hygroscopy	No

*E-mail: ykkim4@hanyang.ac.kr; Fax: +82-2-2296-2354